

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 12 2011

DataMaster cdm S/N 130195

Supervisor/Operator Performing the Verification Procedure:

A Name Robert J. Ely ID# 2681 Date 12-07-2011
Agency Craig Police Department Phone # (907) 826-3330
Instrument Location 506 2nd Street, Craig Alaska 99921

B Alco S/N 75509 Target Value .079 High Pressure 500 PSI
Alco Test Value Average .083 .083
1st Alco 2nd Alco

Signature Robert J. Ely

(OVER)

COB
12/13/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows;

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons Responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of Feb, 2012

Nikki Roth (Notary Seal Stamp)
~~Daniel L. Nicholas~~ Nikki Roth
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

DECEMBER 07, 2011

OPERATOR'S NAME:

ELY, R

OPERATOR'S NUMBER: 2601

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME:

AGE OF SUB:

SEX: M

DOB: 11-11-1988

LAST REPORT: 11-XXXXX

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: 75509

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.76 in
ALCO TARGET .091 12:00
BLANK TEST .000 12:01
INTERNAL STANDARD VERIFIED 12:01
ALCO @ 30.76 in .083 12:01
BLANK TEST .000 12:02
SUBJECT SAMPLE .000 12:03
BLANK TEST .000 12:03
ALCO @ 30.76 in .083 12:04
BLANK TEST .000 12:05

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

DECEMBER 07, 2011

TIME 11:08

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 30.76 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS:
QUARTZ STANDARD:
CALIBRATION: OKAY

INSTRUMENT TEST

11-11-11 11:11:11 11-11-11 11:11:11
11-11-11 11:11:11 11-11-11 11:11:11
11-11-11 11:11:11 11-11-11 11:11:11